

All for Healthcare

NE-C803-AP-09/2014
8200352-0A


	OMRON HEALTHCARE Co., Ltd. 53, Kunotsubo, Terado-cho, Muko, Kyoto, 617-0002 JAPAN
<div>EC</div> <div>REP</div>	OMRON HEALTHCARE EUROPE B.V. Scorpius 33, 2132 LR Hoofddorp, THE NETHERLANDS
Production facility	OMRON HEALTHCARE MANUFACTURING VIETNAM CO., LTD. Binh Duong Province, VIETNAM
Asia Pacific HQ	OMRON HEALTHCARE SINGAPORE PTE LTD. 438A Alexandra Road, #05-05/08 Alexandra Technopark, Singapore 119967 www.omronhealthcare-ap.com

Made in Vietnam

Introduction

Thank you for purchasing Omron Compressor Nebulizer NE-C803. This product was developed in conjunction with respiratory therapists for the successful treatment of asthma, chronic bronchitis, allergies and other respiratory disorders. This is a medical device. Operate the device only as instructed by your doctor and/or respiratory therapist.

How the nebulizer kit works
The medication that is pumped up through the medication channel is mixed with compressed air which is generated by a compressor pump. The medication is turned into fine particles and sprayed when in contact with the compressed air.



Intended use

Medical Purpose
This product is intended to be used for inhaling medication for respiratory disorders.

Intended User
• Legally certified medical experts, such as doctor, nurse and therapist.
• Caregiver or patient under the guidance of qualified medical experts for home treatment.
• The user should also be capable of understanding general operation of NE-C803 and the content of instruction manual.

Intended Patients
This product should not be used on patients, who are unconscious or are not breathing spontaneously.




Environment
This product is intended for use in a medical facility, such as Hospital, clinic and doctor's office, and in a room of general household.

Durable period
Durable periods are as follows, provided the product is used to nebulize 2 ml of medication twice a day for 7.5 minutes each time at room temperature (23°C). Durable period may vary depending on usage environment. Frequent use of the product may shorten the durable period.

Compressor (Main unit)	5 years
Nebulizer Kit	1 year
Mouthpiece	1 year
Nosepiece (optional)	1 year
Air Tube (PVC, 100cm)	1 year
Air Filter	60 days
Mask (PVC)	1 year

Precautions for use
Warnings and cautions described in the instruction manual should be observed.

Know your unit



Important safety instructions

Read all the information in the instruction manual and any other literature included in the box before using the device.

Warning: Indicates a potentially hazardous situation which, if not avoided, could result in serious injury.

(Usage)
• For type, dose, and regime of medication follow the instructions of your doctor or respiratory therapist.
• If you feel anything unusual during use, stop using the device immediately and consult your doctor.
• Do not use only water in the nebulizer for inhaling purposes.
• Keep the device out of reach of unsupervised infants and children. The device may contain small pieces that can be swallowed.
• Do not store the air tube while there is moisture or medication remaining inside.
• Do not use or store the device where it may be exposed to noxious fumes or volatile substances.
• Do not use the device where it may be exposed to flammable gas.
• Do not cover the compressor with a blanket, towel, or any other type of cover during use.
• Always dispose of any remaining medication after use, and use fresh medication each time.
• Do not use in anaesthetic or ventilator breathing circuits.
• Store the parts in a clean location to avoid infection.

(Risk of electrical shock)
• Never plug or unplug the AC adapter with wet hands.
• The compressor and the AC adapter are not waterproof. Do not spill water or other liquids on these parts. If liquid does spill on these parts, immediately unplug the AC adapter and wipe off the liquid with gauze or other absorbent material.
• Do not immerse the compressor in water or other liquid.
• Do not use or store the device in humid locations.
• Do not operate the device with a damaged power cord or plug.
• Keep the power cord away from heated surfaces.
• Use only original OMRON AC adapter. Use of unsupported AC adapter may damage the device.

(Cleaning and disinfecting)
Observe the rules below when cleaning or disinfecting parts. Failure to observe these rules may result in damage, inefficient nebulization or infection. For the instructions, refer to "Cleaning and disinfecting" section.
• Clean and disinfect the nebulizer kit, mask, mouthpiece or optional nosepiece before using them when as follows:
- the first time after purchase.
- if the device has not been used for a long period of time
- if more than one person uses the same device
• Be sure to wash or wipe the parts after use, and ensure that they are thoroughly disinfected and dried, and stored in a clean location.
• Do not leave the cleaning solution in the parts.

Caution: Indicates a potentially hazardous situation which if not avoided, may result in minor or moderate injury, or physical damage.

(Usage)
• Provide close supervision when this device is used by, on, or near children or invalids.
• Make sure that the parts are attached correctly.
• Make sure that the air filter is correctly attached.
• Make sure that the air filter is clean. If the air filter has been used for more than 60 days, replace it with a new one.
• Do not tilt the nebulizer kit at an angle of greater than 45 degrees in all directions or shake it while in use.
• Do not use or store the device while the air tube is creased.
• Use only original nebulizing parts, air tube, air filter and air filter cover.
• Do not add more than 10 ml of medication to the medication tank.
• Do not carry or leave the nebulizer kit while the medication tank contains medication.
• Do not leave the device unattended with infants or persons who cannot express their consent.
• Do not subject the device to any strong shocks such as dropping on the floor.
• Do not jab the nozzle of the medication tank with a pin or any sharp object.
• Do not insert fingers or objects inside the compressor.
• Do not disassemble or attempt to repair the compressor or AC adapter.
• Do not leave the device in extreme hot or cold temperature, or under direct sunlight.
• Do not use the device while sleeping or if drowsy.
• Approved for human use only.
• The compressor may become hot during operation.
• Do not touch the compressor for other than necessary operation such as turning off the power while nebulizing.
• To avoid the medication residue on the face, be sure to wipe the face after removing the mask.
• To avoid injury to the nose mucosa, do not squeeze the optional nosepiece into the back of the nose.
• Do not use a damaged nebulizer kit, mouthpiece or optional nosepiece.

(Risk of electrical shock)
• Always unplug the AC adapter from the power outlet after use and before cleaning.
• Plug the device into the appropriate voltage outlet. Do not overload power outlets or use extension cords.
• Do not misuse the power cord of AC adapter.
• Do not wind the power cord around the compressor or AC adapter.
• Do not pull the power cord of AC adapter.
• Changes or modifications not approved by OMRON HEALTHCARE will void the user warranty.
• Always remove the AC adapter from the device after use and before cleaning.

(Cleaning and disinfecting)
Observe the rules below when cleaning or disinfecting parts. Failure to observe these rules may result in damage, inefficient nebulization or infection. For the instructions, refer to "Cleaning and disinfecting" section.
• Do not use a microwave oven, dish dryer or hair dryer to dry the device or the parts.
• Do not use an autoclave, EOG gas sterilizer or low temperature plasma sterilizer.
• When disinfecting parts by boiling, make sure that the container does not boil dry. Otherwise it may also result in fire.

General Safety Precautions:
• Inspect the device and parts before using them each time, and check that there are no problems. In particular, be sure to check the following:
- That the nozzle or air tube are not damaged.
- That the nozzle is not blocked.
- That the compressor operates normally.
• When using this device, there will be some noise and vibration caused by the pump in the compressor. There will also be some noise caused by the emission of compressed air from the nebulizer kit. This is normal and does not indicate a malfunction.
• Operate the device only as intended. Do not use the device for any other purpose.
• Do not use the device at temperatures greater than +40°C.
• Make sure that the air tube is securely attached to the compressor and nebulizer kit, and does not come loose.
• To completely isolate the device from the power source unplug the plug from the power source.

Keep these instructions for future reference.

Important information regarding Electro Magnetic Compatibility (EMC)
With the increased number of electronic devices such as PC's and mobile (cellular) telephones, medical devices in use may be susceptible to electromagnetic interference from other devices. Electromagnetic interference may result in incorrect operation of the medical device and create a potentially unsafe situation. Medical devices should also not interfere with other devices.

In order to regulate the requirements for EMC (Electro Magnetic Compatibility) with the aim to prevent unsafe product situations, the EN60601-1-2:2007 standard has been implemented. This standard defines the levels of immunity to electromagnetic interferences as well as maximum levels of electromagnetic emissions for medical devices.

This medical device manufactured by OMRON HEALTHCARE conforms to this EN60601-1-2:2007 standard for both immunity and emissions. Nevertheless, special precautions need to be observed:

• Do not use mobile (cellular) telephones and other devices, which generate strong electrical or electromagnetic fields, near the medical device. This may result in incorrect operation of the unit and create a potentially unsafe situation. Recommendation is to keep a minimum distance of 7 m. Verify correct operation of the device in case the distance is shorter.

Further documentation in accordance with EN60601-1-2:2007 is available at OMRON HEALTHCARE EUROPE at the address mentioned in this instruction manual.

Documentation is also available at www.omron-healthcare.com.

Correct Disposal of This Product (Waste Electrical & Electronic Equipment)

This marking shown on the product or its literature, indicates that it should not be disposed of, with other household wastes at the end of its working life. To prevent possible harm to the environment or human health from uncontrolled waste disposal, please separate this product from other types of wastes and recycle it responsibly to promote the sustainable reuse of material resources.

Household users should contact either the retailer where they purchased this product, or their local government office, for details of where and how they can return this item for environmentally safe recycling.

Business users should contact their supplier and check the terms and conditions of the purchase contract. This product should not be mixed with other commercial wastes for disposal.

Accessories (Sold separately)

Name	Model
Nebulizer Kit & Mouthpiece Set	NEB-NSET3-83E
Nebulizer Kit & Adult Mask (PVC) & Air Tube Set	NEB-NSET4-83AP
Nebulizer Kit & Child Mask (PVC) & Air Tube Set	NEB-NSET5-83AP
Adult Mask (PVC)	NEB-MSLP-E
Child Mask (PVC)	NEB-MSMP-E
Mouthpiece	NEB-MP-E
Nosepiece	NEB-NP-E
Air Tube (PVC, 100cm)	NEB-TP-82E
Air Filters x 5	NEB-AFR-30E
AC Adapter	HHP-CF01

Technical data

Name:	COMP AIR
Model:	NE-C803
Type:	Compressor Nebulizer
Rating (AC adapter):	100 - 240V ~ 50/60Hz 0.12 - 0.065A
Rating (Compressor Nebulizer):	6V = 0.7A
Operating Temperature/Humidity:	+10°C to +40°C / 30% to 85%RH
Storage and Transport Temperature/Humidity/ Air Pressure:	-20°C to +60°C / 10% to 95%RH / 700 - 1060hPa
Weight:	Approx. 180g (compressor only)
Dimensions:	Approx. 85 (W) × 43 (H) ×115 (D) mm (compressor only)
Contents:	Compressor, Nebulizer Kit, Air Tube (PVC, 100cm), Mouthpiece, Adult Mask (PVC), Child Mask (PVC), 5 pcs Spare Air Filters, AC Adapter, Storage Bag, Instruction Manual.
Classification:	Class II equipment, Type BF applied part
IP Classification:	IP21
<div> = Class II equipment</div> <div> = Type BF applied part</div> <div> Read the instruction manual carefully</div>	

= No operation

= Power on

0197

Technical data for the OMRON NE-C803 Compressor with the Nebulizer kit:

Particle Size: **MMAD Approx. 3µm
MMAD = Mass Median Aerodynamic Diameter

Medication Tank Capacity: 10ml maximum


Appropriate Medication Quantities: 2ml minimum - 10ml maximum

Nebulization Rate: *Approx. 0.30ml/min (by weight loss)

Aerosol Output: ** 0.37ml (2ml, 1%NaF)

Aerosol Output Rate: ** 0.05ml/min (2ml, 1%NaF)

Result of cascade impactor measurements for particle size**



* Measured by OMRON HEALTHCARE Co., Ltd. Measured with saline at ambient temperature of 23°C and humidity of 40%. It may vary with the type of medication.
** Independently measured by Prof. Dr. Hiroshi Takano, Therapeutic Systems Research Center, Doshisha University, Kyoto, Japan, according to EN 13544-1:2007+A1:2009.

Notes:
• Subject to technical modification without prior notice.
• This OMRON product is produced under the strict quality system of OMRON HEALTHCARE Co., Ltd., Japan.
• The device may not work if the temperature and voltage conditions are different to those defined in the specifications.
• The device fulfils the provisions of the EC directive 93/42/EEC (Medical Device Directive) and the European Standard EN13544-1:2007+A1:2009, Respiratory therapy equipment - Part1: Nebulizing systems and their components.
• Performance may vary with medication such as suspensions or high viscosity. See medication supplier's data sheet for further details.

How to use

⚠ Warning:
Clean and disinfect the nebulizer kit, mask, mouthpiece or optional nosepiece before using them when as follows:


- the first time after purchase.
- if the device has not been used for a long period of time
- if more than one person uses the same device

For the instructions, refer to “Cleaning and disinfecting” section.

1. Make sure that the power switch is in the off (⏻) position.

2. Connect the AC adapter to the compressor and plug the power plug into a power outlet.

3. Remove the inhalation top together with vaporizer head from the medication tank.




Note:
If the vaporizer head came off the inhalation top and fell into the medication tank, remove it out of the tank.

4. Add the correct amount of prescribed medication to the medication tank.




5. Make sure that the vaporizer head is securely attached to the inhalation top.




Note:
The sizes of the left- and right-hand attached parts of the vaporizer head are not the same. If the vaporizer head is not attached in the correct orientation, it will not nebulize correctly.

6. Put the inhalation top back inside the medication tank.

- 1) Align the protrusion on the inhalation top with the indent on the medication tank, as shown.
- 2) Rotate the inhalation top clockwise until the arrow mark on it and the alignment mark on the medication tank are aligned, and they click in place.




7. Attach the mask, mouthpiece or optional nosepiece to the nebulizer kit tightly.




Note:
Make sure that these attachments are fitted in such a way that the handle is positioned away from the patient. Otherwise, medication may be sprayed on patient's skin or clothes.


8. Attach the air tube.
Push the air tube plugs firmly into the air tube connectors.




The nebulizer kit can be fixed on the compressor.



9. Hold the nebulizer kit as indicated below.
Follow the instructions of your doctor or respiratory therapist.




⚠ Caution:
Do not tilt the nebulizer kit at an angle of greater than 45 degrees in all directions. Medication may flow into the mouth or it may result in not nebulizing effectively.



10. Press the power switch to the on (|) position. As the compressor starts, nebulization begins and aerosol is generated.
Inhale the medication.

11. When treatment is complete, turn the power off. Unplug the AC adapter from the power outlet and disconnect it from the compressor.

Note:
Securely hold and press the power plug down to disconnect the AC adapter.



⚠ Caution:
Do not pull the power cord. Damage may be caused.


Removing condensation from the air tube

If there is moisture or liquid remaining in the air tube, be sure to follow the procedure below to remove any moisture from within the air tube.
1) Make sure that the air tube is connected to the air connector on the compressor.
2) Remove the air tube from the nebulizer kit.
3) Turn on the compressor and pump air through the air tube to expel the moisture.


Changing the air filter

If the air filter has been used for more than 60 days, replace it with a new one.

1. Pull the air filter cover off the compressor.



2. Change the air filter.



3. Put the air filter cover back in place.

Make sure it is attached properly.

Notes:
• Use only OMRON air filters designed for this device. Do not operate without air filter.
• Do not wash or clean the air filter. If the air filter becomes wet, replace it. Damp air filters can cause blockages.
• There is no front or back orientation for the air filters.
• Check that air filters are clean and free of dust before inserting them.

Cleaning and disinfecting

⚠ Caution: Handling the Vaporizer Head
• Always wash the vaporizer head after each use.
• Do not use a brush or pin to clean the vaporizer head.
• When disinfecting the parts by boiling, be sure to boil them in plenty of water.
• Do not boil vaporizer head together with objects other than applicable nebulizer accessories.

Cleaning
Clean the parts after each use to remove residual medication. This will avoid inefficient nebulization or infection.

■ Washable parts

• **Nebulizer Kit (Inhalation top, Medication tank), Mask (PVC), Mouthpiece, Nosepiece (optional)**
Wash them in warm water and mild detergent (neutral detergent). Rinse them thoroughly with clean hot tap water and allow to air dry in a clean place.
• **Vaporizer Head**
Wash it with running water.

■ Non-washable parts

• **Compressor, Air Tube**
Firstly, make sure that the power plug is unplugged from the power outlet. Wipe clean with a soft cloth moistened with water or mild detergent (neutral detergent).
• **Air Filter**
Do not wash or clean the air filter. If the air filter becomes wet, replace it. Damp air filters can cause blockages.

Disinfecting
Disinfect the parts once a week. If the parts are heavily stained, replace them with new ones.

To select a method for disinfection, refer to the table below.


○: applicable X: not applicable

Disinfecting method	Parts	Mouthpiece	Nosepiece (optional)	Nebulizer Kit	Adult Mask (PVC) Child Mask (PVC)	Air Tube (PVC, 100cm)	Air Filter
	Materials	PP	PP	PP (Vaporizer Head: PC)	Mask: PVC (Phthalate free) Band: Rubber (non-latex)	PVC (Phthalate free)	Polyester
A	Alcohol	○	○	○	○	X	X
	Disinfecting ethanol						
	Sodium hypochlorite	○	○	○	○	X	X
	Milton* (0.1%, 15min.)						
	Quaternary ammonium	○	○	○	○	X	X
	Osvan* (0.1%, 10min.)						
	Chlorhexidine	○	○	○	○	X	X
B	Hibitane* (0.5%, 30min.)						
	Amphoteric Surfactant	○	○	○	○	X	X
	Tego 51* (0.2%, 15min.)						
B	Boiling	○	○	○	X	X	X

* an example of commercially available disinfectant. (The concentration and residence time specified in the table are under the conditions where the service life of the parts is tested with each disinfectant used as described in its instruction manual. Please note that the testing was not performed for the purpose of ensuring the effectiveness of the disinfectants. There is no intention of suggesting that such disinfectants be used. Conditions of use and ingredients of disinfectant vary for manufacturers. Please read the instruction manuals carefully before use and apply disinfectant to each part in an appropriate way. Please notice that service life of the parts may be shorter depending on conditions, environments and frequency of use.)

A. Use a commercially available disinfectant. Follow the instructions provided by the disinfectant manufacturer.
Note: Never clean with benzene, thinner or a flammable chemical.

B. Parts may be boiled between 15 to 20 minutes.
After boiling, carefully remove the parts, shake off excess water and allow to air dry in a clean environment.



Troubleshooting

Check the following if the device should fail during operation.

Problem	Cause	Remedy
Nothing happens when the power switch is pressed.	Is the AC adapter connected correctly to a power outlet and the compressor?	Check that the AC adapter is connected to a power outlet and the compressor. Unplug then reinsert the plug if necessary.
No nebulization or low nebulization rate, when the power is on.	Is there medication in the medication tank?	Add the correct amount of medication to the medication tank.
	Is there too much/little medication in the medication tank?	
	Is the vaporizer head missing or not assembled correctly?	Attach the vaporizer head correctly.
	Is the nebulizer kit assembled correctly?	Assemble the nebulizer kit correctly.
	Is the nozzle blocked?	Make sure that the nozzle is free of blockages.
	Is the nebulizer kit tilted at a sharp angle?	Make sure that the nebulizer kit is not tilted at an angle of more than 45 degrees.
	Is the air tube connected correctly?	Make sure that the air tube is correctly connected to the compressor and nebulizer kit.
	Is the air tube folded?	Make sure that the air tube is not twisted or bended.
	Is the air tube blocked?	Make sure that the air tube is free of blockages.
The compressor is abnormally loud.	Is the air filter dirty?	Replace the air filter with a new one.
	Is the air filter cover attached correctly?	Attach the air filter cover correctly.
The compressor is very hot.	Is the compressor being covered?	Do not cover the compressor during operation.

Note: If the suggested remedy does not solve the problem, do not try to repair the device - the device are not user serviceable. Please contact an authorized OMRON representative for assistance.